



NATIONAL INSTITUTE OF TECHNOLOGY SRINAGAR

(An Autonomous Institute of National Importance Established by the Act of Parliament)

OFFICE OF THE COORDINATOR TEQIP III

Tel: +91-194-2422032 Extn: 2818, 2814, 2806

Email: teqip3@nitsri.net Website: <http://new.nitsri.ac.in>


No. NIT/TEQIP/19/123

Dated: 18-02-2020

Corrigendum

With reference to Invitation for quotation letter vide no NIT/TEQIP20/112 Dated 12-02-2020, kindly read specification in **Annexure I** as:

S No.	ITEM NAME	SPECIFICATIONS	QUANTITY
1	Sonicator	2.5 L Tank Capacity Ultrasonic Frequency 30-40 kHz Time Setting 1- 30 minutes Temperature 80 degree Celsius Max	1
2	Centrifuge (Benchtop)	Variable speed 6000rpm; 2000g 8 tubes(0.2/0.5/1.5/2 ml) Electronic Brake for immediate stop on lid opening.	1
3	Variable Micropipette	Manual Single Channel Pipette. Volume Range 20-200 μ L with increment 1.0 μ L	1
4	Variable Micropipette	Manual Single Channel Pipette. Volume Range 2-20 μ L with increment 0.1 μ L	1


Nodal officer Procurement
TEQIP III